

Instruction for Use

PCN Catheter

Device Description

The PCN Catheters are provided for use in Urology procedures for temporary urinary diversion associated with urinary obstruction secondary to calculi. Intended for single use. The PCN Catheter consists of a Radiopaque pigtail catheter and a Urine Bag connector. The configurations available include:

| Family | Accessories | Size (Fr.) | Length (cm) | Catalogue Number | Color |
|--------------------|-------------------------------------|------------|-------------|------------------|-------|
| PCN Catheter (PCN) | Straightener Urine Bag Connector | 6 | 22 30 | AMPLCT004 | Blue |
| | | 7 | | | |
| | | 8 | | | |
| | | 9 | | | |
| | | 10 | | | |
| | | 12 | | | |
| | | 14 | | | |

Intended Purpose

PCN Catheters used for percutaneous placement of a pigtail catheter in the renal pelvis for nephrostomy drainage.

Performance Characteristics of the device

Percutaneous Nephrostomy (PCN) tube is a catheter that is inserted through your skin into the kidney.

The nephrostomy tube placed is sterile and radiopaque for better visualization. The pigtail tip configuration aids in the retention of the distal tip in the renal pelvis to drain urine from the body into a collecting bag outside the body.

Indications

- Used in patients diagnosed with large kidney Stones, blocking more than one branch of the collecting system of the kidney (known as staghorn kidney stones).
- Presence of larger stones in ureter.
- Uretero-Pelvic Junction Obstruction.
- Decompression of the renal collecting system.

Contraindications

- Bleeding Diathesis (most commonly uncontrollable coagulopathy or pharmacological anticoagulation).
- Severe hyperkalemia
- Severe metabolic acidosis
- Uncooperative patient.

Precautions & Warnings

Carefully read all instructions for use and product labeling. Do not use this product without reading and understanding the complete instructions enclosed herein. The device shall only be applied for its intended use and in accordance with these instructions. Observe all cautions and warnings throughout these instructions. Failure to do so may result in complications.

Precautions:

- The intended user of the device must be a Urologist specializing in the treatment of urinary system disorders and related urological procedures authorized by the Competent Authority of the country in which the Physician is practicing.
- Each Physician is responsible for using the appropriate technique and deciding on the indication for use of this device.
- Confirm the information on the label and that the product has not reached its expiration date and there is no damage to the packaging or device.
- Device is not recommended for use in patients with the above-mentioned contraindicated conditions.
- Store the PCN Catheter at a temperature between 12-35°C and a Humidity range of 30-75%.
- Store in a dry, cool place. Avoid extended exposure to sun light.

Warnings:

- All components of the PCN Catheters/ Nephrostomy Catheter are for single use only.
- The product must not be re-used. Reusing single-use devices can lead to potentially serious consequences for the patient such as bio-contamination due to release of infectious agents from device into the body which further may result in Urinary Infection.
- Do not use the device if there is any indication that the sterility of the device has been compromised. If the sterile package is damaged or possibly opened, do not use. Contact "Manufacturer or Distributor" and replace the product.
- Do not reprocess or re-sterilize, reprocessing or re-sterilization may compromise the structural integrity of the device of the device and/or lead to failure which, in turn, may result in patient injury.

Intended Patient Population

The device is intended for use in patients of all age groups.

Shelf-life

The shelf life of the PCN Catheters is 3 years (Indicated on product label with the following use-by date symbol).

Sterility – This product is Sterile unless the package has been opened or damaged

- The PCN Catheters have been sterilized by exposure to Ethylene Oxide. Sterility indicators are on each package. The imprinted label will change color from blue to brown after ethylene oxide exposure. Do not use the product unless the sterility indicators are the correct color.

Instruction for Use

PCN Catheter

- Exposure to high levels of Ethylene Oxide may result in presence of residual ETO which leads to Toxic Reaction in the body resulting in tissue damage of Urinary System.
- The PCN Catheters are packed single. Product in each pack must be utilized immediately when opened.

Directions for Use

Open the package from the Chevron Side ('V' Side) of the pouch.

- By preliminary plain film, I.V.P, retrograde pyelogram, ultrasound or CT scan, localize the kidney to be drained. A prone or slightly oblique position is preferred. Prepare and drape the flank in the usual fashion. Under fluoroscopic control, identify and anesthetize the skin site overlying the collection system.
- Pass the 18 gauge needle vertically through a small incision in the skin (made with a scalpel blade) insert the needle into the appropriate part of the pyelocaliceal system. Free flow of fluid from the needle after removing the obturator confirms a satisfactory location for the tip of the needle cannula.
- Pass the flexible end of the J guide wire through the hub of the 18-gauge needle cannula, insert the guide wire three (3) more inches into the collecting system. Confirm the position fluoroscopically. Maintain the position of the guide wire as the needle cannula is withdrawn over it.
- Care must be exercised to avoid withdrawal of the guide wire itself
- To facilitate passage of the PCN catheter, dilate the musculofascial tract by progressing from the smallest to the largest dilator provided.
- Pass the pigtail end of the catheter over the external end of the guide wire; gradually advance the pigtail end well into the collecting system. Confirm the position fluoroscopically. While the shaft of the pigtail catheter is held securely in position with one hand, the guide wire is withdrawn with the other.
- Use the urine bag connector to connect the pigtail catheter to a drainage bag or leg bag.






















Removal Instructions

- Retrieve cystoscopically by gently pulling on Catheter or retrieval line with grasping forceps or equivalent.
- If resistance is encountered during removal of the Catheter, stop and determine cause of resistance before proceeding.

Disposal Instructions

Dispose of all equipment, in appropriate containers. After use, this product may be a potential biohazard. Handle and dispose of in accordance with accepted medical practice and applicable local, state and federal laws and regulations.

Explanation of symbols used on label

| Symbol | Title of Symbol | Symbol | Title of Symbol | Symbol | Title of Symbol | Symbol | Title of Symbol | Symbol | Title of Symbol |
|--|---------------------------------|---|------------------------------|---|-------------------------|---|----------------------------------|---|--|
|  | Catalogue Number |  | CE Mark |  | Do not Re sterilize |  | Country of Manufacture |  | Medical Device |
|  | Manufacturer |  | Date of manufacture |  | Keep Dry |  | Caution |  | Single Sterile barrier system with protective packaging inside |
|  | Batch Code |  | Use by Date |  | Keep Away from Sunlight |  | Do not re-use |  | Unique Device Identifier |
|  | Sterilized using ethylene oxide |  | Consult Instructions for Use |  | Temperature Limit |  | Do Not Use if Package is Damaged |  | Authorized Representative in the European community |
|  | Humidity Limit | | | | | | | | |



Aster Medispro Pvt. Ltd
 S.P.181, 10th Main, 1st Stage,
 DR.B.R.Ambedkar Industrial Estate (KSSIDC)
 Jigani Industrial area, Jigani,
 Bangalore- 560105, Karnataka, India.
 Tel: +91 80-42062716
 Email: info@astermedispro.net
 Web: www.astermedispro.net



M/sCMC Medical Devices & Drugs S.L.
 located in C/Horacio Lengo N° 18, CP29006,
 Málaga, Spain
 Tel : +34951214054, Fax: +34952330100
 E-mail : mmateos@cmcmedicaldevices.com
 Web : www.cmcmedicaldevices.com

Bibliography

- Contraindications: <https://www.ncbi.nlm.nih.gov/pmc/articles/PMC3312169/>
- Single Use Medical Device:
https://assets.publishing.service.gov.uk/government/uploads/system/uploads/attachment_data/file/956268/Single_use_medical_devices.pdf
- EN ISO 15223-1:2021 – Medical devices – Symbols to be used with medical device labels, labelling and information to be supplied – Part 1: General Requirements
- EN ISO 20417:2020 – Medical Devices – Information to be supplied by the manufacturer.